## **VACTEST PPP 700 D**

Measuring device for food packaging processes





## **Packageable**

Specially designed for direct use in food packaging processes, can be placed inside the packaging and be packaged, robust and compact construction

## **High performance**

Excellent measurement accuracy due to Piezo-Pirani combination sensor, great measurement range (1,200 - 0.01 mbar)

## **Intuitive operation and handling**

Easy control via five buttons, easy-to-read display, integrated li-ion battery, rechargeable via USB-C

www.buschvacuum.com © Busch Vacuum Solutions 13.02.2024 1/2

# **VACTEST PPP 700 D**

Measuring device for food packaging processes



|                             | VACTEST PPP 700 D   |
|-----------------------------|---|
| Measurement principle       | Piezoresistive / Pirani   |
| Measurement range           | 1200-0.01 mbar  |
| Resolution                  | 1200–1 mbar: 0.1 mbar , 0.99–0.01 mbar: 0.01 mbar                     |
| Measurement uncertainty     | 1200–40 mbar: 0.3% full scale , 40–0.01 mbar: 10% of reading          |
| Measurement interval        | 0.02-60 s   |
| Supply voltage              | Internal rechargeable lithium-ion battery                             |
| Min. battery operating time | 48 h  |
| Serial interface            | USB-C and Bluetooth   |
| Display                     | LCD graphic display, resolution 400 x 240 px                          |
| Operating temperature       | +5 +50 °C   |
| Weight approx.              | 170 g   |
| Dimensions (L x W x H)      | 105 x 70 x 25.4 mm  |
| Vacuum connection           | Sensor flush mounted in housing, G 1/8" female thread for calibration |

## **DO YOU WANT TO KNOW MORE?**

Get in touch with us directly! busch@busch.at or +43 (0)2262 75665-0





CALL NOW

www.buschvacuum.com © Busch Vacuum Solutions 13.02.2024 2/2